Instrument Compatibility

One glance at the table will tell you which BRAND micro-, deep well- and PCR-plates are compatible with your instruments like dispensers, plate reader, pipetting robots, etc.
We are continuously updating the table with information from manufacturers and feedback from our customers.
Ask for a free sample of our plates to check compatibility with your instrument (www.brand.de), with no obligation.
A short feedback from you about the result will help us and other customers to complete the table.

A regularly updated table can be found at www.brand.de

	96-well BRAND*plates*®											
	Standard										with transp. Bottom	
	U	V	F	C	U	F	C	U	F	C	F	F
	transparent	transparent	transparent	transparent	white	white	white	black	black	black	white	black
	781600 / 781660 781720 / 781780 / 781840 781900 / 781960	781601 / 781661 781721 / -1841 781961	781602 / 781662 781722 / 781782 / 781842 781962/782022/782082/781900	781603 / 781663 781723 / 781783 / 781843 781963	781604 781724 /781784 / 781844 781904	781605 / 781665 781965/782025/782085/781905	781726 / 781786 / 781846 781906	781607 781727 / 781787 / 781847 781907	781608 / 781668 781968/782028/782088/781908	781729 / 781789 / 781849 781909	781610 / 781670 781850 781970/782030/782090/781910	781611 / 781671 781731 / 781851 781971/782031/782091/781911
Dispenser												
BioTek Instruments GmbH												
MicroFill™	•	•	•	•	•	•	•	•	•	•	•	•
MultiFlo™	•	•	•	•	•	•	•	•	•	•	•	•
EL406™	•	•	•	•	•	•	•	•	•	•	•	•
HCS-Reader												
innovatis AG												
Cellavista Analyzer	–	–	•	•	–	–	–	–	–	–	•	•
Microfluidics												
Caliper LifeSciences GmbH												
LabChip® 3000												
Plate incubator												
BMG LABTECH GmbH												
THERMOstar	•	•	•	•	•	•	•	•	•	•	•	•
Plate reader												
Analytik Jena AG												
FLASHScan	•2)	•2)	•	•	•2)	•2)	•2)	•2)	•2)	•2)	•	•
SpeedScan	•	•	•	•	•	•	•	•	•	•	•	•
Berthold Detection Systems GmbH												
MLP1	–	–	–	–	•	•	•	•1)	•1)	•1)	•	•1)
MLP2	–	–	–	–	•	•	•	•1)	•1)	•1)	•	•1)
Orion II	–	–	–	–	•	•	•	•1)	•1)	•1)	•	•1)
Orion L	–	–	–	–	•	•	•	•1)	•1)	•1)	•	•1)
BERTHOLD TECHNOLOGIES GmbH & Co.KG												
Apollo-1 LB 911					–	–	–	–	–	–	•	•
Apollo-8 LB 912	•	•	•	•	–	–	–	–	–	–	•	•
Centro / Centro XS3 LB 960	–	–	–	–	•	•	•	–	–	–	•	–
CentroLIA LB 961	–	–	–	–	•	•	•	–	–	–	–4)	–4)
CentroLIA*pc* LB 962	–	–	–	–	•	•	•	–	–	–	–4)	–4)
Mithras LB 940	•	•	•	•	•	•	•	•	•	•	•	•
TriStar LB 941	•	•	•	•	•	•	•	•	•	•	•	•
Twinkle LB 970	–	–	–	–	–	–	–	•	•	•	–	•
BioTek Instruments GmbH												
Eon™	•	•	•	•	–	–	–	–	–	–	•1)	•1)
Epoch™	•	•	•	•	–	–	–	–	–	–	•1)	•1)
PowerWaveHT™	•	•	•	•	–	–	–	–	–	–	•1)	•1)
ELx800™	•	•	•	•	–	–	–	–	–	–	•1)	•1)
ELx808™	•	•	•	•	–	–	–	–	–	–	•1)	•1)
FLx800™	–	–	–	–	•	•	•	•	•	•	•	•
Synergy™ H4	•	•	•	•	•	•	•	•	•	•	•	•
Synergy™ H1	•	•	•	•	•	•	•	•	•	•	•	•
Synergy™ Mx	•	•	•	•	•	•	•	•	•	•	•	•
Synergy™ 2	•	•	•	•	•	•	•	•	•	•	•	•
Synergy™ HT	•	•	•	•	•	•	•	•	•	•	•	•
BMG LABTECH GmbH												
FLUOstar Omega	•	•	•	•	•	•	•	•	•	•	•	•
FLUOstar OPTIMA	•	•	•	•	•	•	•	•	•	•	•	•
LUMIstar Omega	–	–	–	–	•	•	•	–	–	–	•	–
LUMIstar OPTIMA	–	–	–	–	•	•	•	–	–	–	•	–
NEPHELOstar												
NOVOstar	•	•	•	•	•	•	•	•	•	•	•	•
PHERAstar	•	•	•	•	•	•	•	•	•	•	•	•
POLARstar Omega	•	•	•	•	•	•	•	•	•	•	•	•
POLARstar OPTIMA	•	•	•	•	•	•	•	•	•	•	•	•
RUBYstar	–	–	–	–	–	–	–	•	•	•	–	–
Spectrostar Omega	•	•	•	•	–	–	–	–	–	–	•	•
STRATEC Biomedical Systems AG												
Photuris												

	384-well BRAND*plates*®								
	Standard			Std., Low Volume			Standard, UV	with transp. Bottom	
	F	V	F	F	V	F	F		
	transparent	white	black	transparent	white	black	UV-transparent	white	black
	781620 / 781680 781740 / 781800 / 781860 781980/782040/782100/781920	781621 / 781681 781741 / 781801 / 781861 781981 / 781921	781622 / 781682 781742 / 781802 / 781862 781982 / 781922	781623 / 781683 781863 781983	781624 / 781684 781864 781984	781625 / 781685 781865 781985	781628 / 781688	781626 / 781686 781866 781986/782046/782106/781926	781627 / 781687 781867 781987/782047/782107/781927
Dispenser									
BioTek Instruments GmbH									
MicroFill™	•	•	•	•	•	•	•	•	•
MultiFlo™	•	•	•	•	•	•	•	•	•
EL406™	•	•	•	•	•	•	•	•	•
HCS-Reader									
innovatis AG									
Cellavista Analyzer	•	–	–	–	–	–	•	•	•
Microfluidics									
Caliper LifeSciences GmbH									
LabChip® 3000				•	•	•			
Plate incubator									
BMG LABTECH GmbH									
THERMOstar	•	•	•	•	•	•	•	•	•
Plate reader									
Analytik Jena AG									
FLASHScan	•	•2)	•2)	•	•2)	•2)	•	•	•
SpeedScan	•	•	•	•	•	•	•	•	•
Berthold Detection Systems GmbH									
MLP1	–	–	–	–	–	–	–	–	–
MLP2	–	–	–	–	–	–	–	–	–
Orion II	–	•	•1)	–	–4)	–4)	–	–4)	–4)
Orion L	–	–	–	–	–	–	–	–	–
BERTHOLD TECHNOLOGIES GmbH & Co.KG									
Apollo-1 LB 911	•	–	–	•	–	–	•	–	–
Apollo-8 LB 912	–	–	–	–	–	–	–	–	–
Centro / Centro XS3 LB 960	–	–	–	–	•	–	–	•	–
CentroLIA LB 961	–	–	–	–	–	–	–	–	–
CentroLIA*pc* LB 962	–	–	–	–	–	–	–	–	–
Mithras LB 940	•	•	•	•	•	•	•	•	•
TriStar LB 941	–	–	–	•	•	•	–	•	•
Twinkle LB 970	–	–	•	–	–	•	–	–	•
BioTek Instruments GmbH									
Eon™	•	–	–	•	–	–	•	•1)	•1)
Epoch™	•	–	–	•	–	–	•	•1)	•1)
PowerWaveHT™	•	–	–	•	–	–	•	•1)	•1)
ELx800™	•	–	–	•	–	–	•	•1)	•1)
ELx808™	–	–	–	–	–	–	–	–	–
FLx800™	–	•	•	–	•	•	–	•	•
Synergy™ H4	•	•	•	•	•	•	•	•	•
Synergy™ H1	•	•	•	•	•	•	•	•	•
Synergy™ Mx	•	•	•	•	•	•	•	•	•
Synergy™ 2	•	•	•	•	•	•	•	•	•
Synergy™ HT	•	•	•	•	•	•	•	•	•
BMG LABTECH GmbH									
FLUOstar Omega	•	•	•	•	•	•	•	•	•
FLUOstar OPTIMA	•	•	•	•	•	•	•	•	•
LUMIstar Omega	–	•	–	–	•	–	–	•	–
LUMIstar OPTIMA	–	•	–	–	•	–	–	•	–
NEPHELOstar		–	–		–	–			
NOVOstar	•	•	•	•	•	•	•	•	•
PHERAstar	•	•	•	•	•	•	•	•	•
POLARstar Omega	•	•	•	•	•	•	•	•	•
POLARstar OPTIMA	•	•	•	•	•	•	•	•	•
RUBYstar	–	•	•	–	•	•	–	•	•
Spectrostar Omega	•	–	–	•	–	–	•	•	•
STRATEC Biomedical Systems AG									
Photuris	•	•	•	•	•	•	•	•	•

	1536-well BRANDplates®			Deep-Well Platten						PCR-Platten						
	Standard															
	F	V	F													
	transparent	white	black	96-well, PP, 0,3 ml	96-well, PP, 0,5 ml	96-well, PP, 1,1 ml	96-well, PS, 1,1 ml	96-well, PP, 2,2 ml	384-well, PP, 0,3 ml	96-well, ohne R.	96-well, erhöhter R.	96-well, ganzer R.	96-well, halber R.	384-well, ganzer R.	384-well, ganzer R.	384-well, ganzer R.
	781640 / 781700 782000	781641 / 781701 782001	781642 / 781702 782002	701330	701346	701350	701352	701354	701355	781350	781352	781353	781400	781345	781347	781348
Dispenser																
BioTek Instruments GmbH																
MicroFill™	–	–	–	•	•	•	•	•	•	•	•	•	•	•	•	•
MultiFlo™	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•	•
EL406™	•	•	•	–	–	–	–	–	–	•4)	•	•	•4)	•	•	•
HCS-Reader																
innovatis AG																
Cellavista Analyzer	•	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
Microfluidics																
Caliper LifeSciences GmbH																
LabChip® 3000	–	–	–													
Plate incubator																
BMG LABTECH GmbH																
THERMOstar	•	•	•													
Plate reader																
Analytik Jena AG																
FLASHScan	•	•2)	•2)	•2)						•2)	•2)	•2)	•2)	•2)	•2)	•2)
SpeedScan	–	–	–	•						•	•	•	•	•	•	•
Berthold Detection Systems GmbH																
MLP1	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
MLP2	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
Orion II	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
Orion L	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
BERTHOLD TECHNOLOGIES GmbH & Co.KG																
Apollo-1 LB 911	–	–	–													
Apollo-8 LB 912	–	–	–													
Centro / Centro XS3 LB 960	–	–	–													
CentroLIA LB 961	–	–	–													
CentroLIA*pc* LB 962	–	–	–													
Mithras LB 940	•	•	•													
TriStar LB 941	–	–	–													
Twinkle LB 970	–	–	–													
BioTek Instruments GmbH																
Eon™	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
Epoch™	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
PowerWaveHT™	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
ELx800™	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–	–
ELx808™				–	–	–	–	–	–	–	–	–	–	–	–	–
FLx800™	–	–	–	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
Synergy™ H4	•	–	•	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
Synergy™ H1	–	–	–	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
Synergy™ Mx	–	–	–	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
Synergy™ 2	•	–	•	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
Synergy™ HT	–	–	–	–	–	–	–	–	–	•3)	•	•	•3)2)	□	□	□
BMG LABTECH GmbH																
FLUOstar Omega	–	–	•													
FLUOstar OPTIMA	–	–	•													
LUMIstar Omega	–	–	–													
LUMIstar OPTIMA	–	–	–													
NEPHELOstar	–	–	–													
NOVOstar	–	–	–													
PHERAstar	•	•	•													
POLARstar Omega	–	–	•													
POLARstar OPTIMA	–	–	•													
RUBYstar	–	–	–													
Spectrostar Omega	–	–	–													
STRATEC Biomedical Systems AG																
Photuris	•	•	•													

□ = not tested yet, • = suitable, •1) = not used often, •2) = only Top Reading, •3) = with adaptor, - = not compatible, -4) = not recommended

Instrument Compatibility

One glance at the table will tell you which BRAND micro-, deep well- and PCR-plates are compatible with your instruments like dispensers, plate reader, pipetting robots, etc.
We are continuously updating the table with information from manufacturers and feedback from our customers.
Ask for a free sample of our plates to check compatibility with your instrument (www.brand.de), with no obligation.
A short feedback from you about the result will help us and other customers to complete the table.

A regularly updated table can be found at www.brand.de

	96-well BRANDplates®												384-well BRANDplates®												1536-well BRANDplates®			Deep-Well Platten						PCR-Platten						
	Standard	U	V	F	C	U	F	C	U	F	C	with transp. Bottom	F	F	Standard	F	V	F	Std., Low Volume	F	V	F	Standard, UV	UV-transparent	with transp. Bottom	F	F	Standard	F	V	F	96-well, PP, 0,3 ml	96-well, PP, 0,5 ml	96-well, PP, 1,1 ml	96-well, PS, 1,1 ml	96-well, PP, 2,2 ml	384-well, PP, 0,3 ml	96-well, ohne R.	96-well, erhöhter R.	96-well, ganzer R.	96-well, halber R.	384-well ganzer R.	384-well, ganzer R.	384-well, ganzer R.
		transparent	transparent	transparent	transparent	white	white	white	black	black	black		white	black		transparent	transparent	white		transparent	white	black				white	black		transparent	white	black												
		781600 / 781660	781720 / 781760 / 781840	781000 / 781040	781001 / 781041	781721 / 1841	781661	781662	781722 / 781762 / 781842	781962/9.302222/ 630562/ 781900	781663	781723 / 781763 / 781843	781963	781604	781724 / 781784 / 781844	781004	781605 / 781665	781965/9.302557/630565/781905	781726 / 781766 / 781846	781906	781607	781727 / 781767 / 781847	781907	781606 / 781668	781968/9.302567/630568/781908	781729 / 781769 / 781849	781909	781610 / 781670	781850	781910	781611 / 781671	781731 / 781851	781911	781620 / 781680	781407 / 781900 / 781890	781940/9.304072/630100/781920	781621 / 781681	781441 / 781901 / 781961	781921	781622 / 781682	781442 / 781902 / 781962	781922	781623 / 781683	781863	781983	781624 / 781684	781864	781984	781625 / 781685	781865	781985	781626 / 781686	781866	781986	781943/9.304077/630107/781927	781627 / 781687	781967	781944/9.304177/630107/781927	781640 / 781700	782000	781641 / 781701	782001	781642 / 781702	782002	701330	701346	701390	701352	701354	701355	781350	781352	781353	781400	781345	781347	781348
Pipetting robots																																										
Analytik Jena AG																																									
FasTrans		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●	●	●	●	●	●	●	●	●	●	●	–	●	●	●	●	●
Beckman Coulter GmbH																																										
Biomek 2000		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
Biomek 3000		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
Biomek FX		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●	●	●												
Biomek NX		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●	●	●												
Multimek		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●	●	●												
BioTek Instruments GmbH																																										
Precision™		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
Caliper LifeSciences GmbH																																										
Sciclone		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●3)	●3)	●3)												
Zephyr		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●3)	●3)	●3)												
ltf labortechnik GmbH & Co KG																																										
CAS-1200		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●1)	●1)	●1)												
CAS-4200		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
CAS-4800		●	●	●	●	●	●	●	●	●	●	●	●		–	–	–	–		–	–	–		–	–			–	–	–												
STRATEC Biomedical Systems AG																																									
Reflex															●	●	●	●		●	●	●		●	●			–	–	–												
Washer																																										
BioTek Instruments GmbH																																										
ELx405™		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
ELx50™		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			–	–	–												
EL406™		●	●	●	●	●	●	●	●	●	●	●	●		●	●	●	●		●	●	●		●	●			●	●	●												

Our technical literature is intended to inform and advise our customers. However, the validity of general empirical values, and of results obtained under test conditions, for specific applications depend on many factors beyond our control. Please appreciate, therefore, that no claims can be derived from our advice. The user is responsible for checking the appropriateness of the product for any particular application.
Subject to technical modification without notice. Errors excepted.

BRAND GMBH + CO KG · P.O. Box 1155 · 97861 Wertheim · Germany · Tel.: +49 9342 808-0 · Fax: +49 9342 808-98000 · E-Mail: info@brand.de · Internet: www.brand.de

□ = not tested yet, ● = suitable, ●1) = not used often, ●2) = only Top Reading, ●3) = with adaptor, - = not compatible, -4) = not recommended