

Transferpette® S



Genuine one-hand operation -
set and lock volume, pipet, eject

Distinct precision and
effortless operation

Volume-change protection
for safe pipetting



Transferpette® S single and multichannel pipette

Precise and easy to use – the Transferpette® S microliter pipettes provide all features required for pipetting in Life Science applications! By using innovative materials, the Transferpette® S pipette has become a perfect manual pipette for demanding laboratory applications.

- Effortless one-handed operation – set and locking the volume, pipetting, ejecting – all without shifting hand position!
- Clearly visible volume display – 4 digits for precise setting. Always visible, even while pipetting.
- Completely autoclavable at 121 °C (20 min) without disassembly.
- Calibration adjustment with the Easy Calibration technique.

Transferpette® S – the result of intensive ergonomic and operational studies

The genuine one-handed operation including volume adjustment for right and left-handers, as well as the reduced weight ensure easy and efficient pipetting in every volume range.

The short 12.5 mm stroke reduces the risk of RSI (Repetitive Strain Injury). This makes the Transferpette® S a very pleasant, reliable pipette with respect to its features!

The particular advantage of the multichannel pipette is easy operation, e.g., when performing serial pipetting of immunological assays, while making serial dilutions, or when filling 96-well plates for cell cultures. The perfect complement to the Transferpette® S single-channel pipette!



More features

- Large, central pipetting button and separate ejection function
- Volume-change protection
- Corrosion-resistant pistons and ejector
- Color-coded volume range
- Transferpette® S 0.1 - 1 µl – highest precision for small volumes in molecular biology, especially for pipetting enzymes.
- 3-year warranty for single-channel pipettes
- UV resistant
- CE-**IVD**-compliant

Transferpette® S

Volume range:

The Transferpette® S pipette provides maximum versatility and optimum quality over the entire volume range.

There are 10 adjustable-volume single-channel pipettes and 9 fixed-volume single-channel pipettes available in the 0.1 µl to 10 ml range.

Transferpette® S-8



There are 5 sizes of multichannel pipettes available in the 0.5 to 300 µl volume range.

A perfect team with pipette tips by BRAND

BRAND pipette tips and filter tips fit perfectly on the Transferpette® S and many other pipettes.

The tips are manufactured from high-quality materials in a state-of-the-art cleanroom in Germany. Thus, all racked tips and

filter tips up to 1000 µl are free of DNA (<40 fg), RNase (<8.6 fg), endotoxins (<1 pg), and ATP (<1 fg), and ensure contamination-free work.

BRAND offers a full tip system for any application requirement.



At a glance ...



Genuine one-hand operation



Label on the finger rest for individual marking



Changes from factory calibration settings are externally visible by recalibration flag

Centrally located!
Large pipetting button allows uniform, smooth movement of the piston.

Secure!
A brake helps protect against accidental volume changes.

Clearly accurate!
4-digit volume display for precise volumes. Always remains visible during pipetting.

Color-coded!
Color indicates volume range

Slim!
If the tip fits in the vessel, so does the shaft – no retrofitting necessary!

Maximum compatibility!
The tip cone design offers excellent fit with pipette tips from BRAND and broadest compatibility with tips from most other manufacturers.



Fast!
One-handed volume setting
for both right- and left-
handed users – even with
gloves!

Effortless!
Ergonomically arranged
tip ejection button with
color code according to
volume range.

Maximum protection!
Completely autoclavable
at 121 °C (20 min) without
disassembly – for maximum
protection from contamination!

Low strain!
Special O-rings made of
silicone / FKM and the stepped
design of the tip ejector reduce
attachment and ejection forces.

Individual shafts and
seals of the multichannel
pipette Transferpette® S
-8/-12 can be easily
unscrewed (patented)
and easily cleaned or
replaced. Saves cost
and downtime!



Technical Data

Items supplied:

Transferpette® S, adjustable volume, DE-M marking, performance certificate, shelf/rack mount and silicone grease.



Transferpette® S, adjustable volume

| Capacity, µl (color-coded) | Description | A* ≤ ± | | CV* ≤ | | Subdivision µl | Cat. No. |
|-------------------------------|-------------|--------|-------|-------|-------|-------------------|----------------|
| | | % | µl | % | µl | | |
| ● 0.1 - 1 | D-1 ① | 2 | 0.02 | 1.2 | 0.012 | 0.001 | 7047 68 |
| ● 0.1 - 2.5 | D-2,5 ② | 1.4 | 0.035 | 0.7 | 0.018 | 0.002 | 7047 69 |
| ● 0.5 - 10 | D-10 | 1 | 0.1 | 0.5 | 0.05 | 0.01 | 7047 70 |
| ● 2 - 20 | D-20 | 0.8 | 0.16 | 0.4 | 0.08 | 0.02 | 7047 72 |
| ● 5 - 50 | D-50 | 0.8 | 0.4 | 0.3 | 0.15 | 0.05 | 7047 73 |
| ● 10 - 100 | D-100 | 0.6 | 0.6 | 0.2 | 0.2 | 0.1 | 7047 74 |
| ● 20 - 200 | D-200 | 0.6 | 1.2 | 0.2 | 0.4 | 0.2 | 7047 78 |
| ● 100 - 1000 | D-1000 | 0.6 | 6 | 0.2 | 2 | 1 | 7047 80 |
| ● 500 - 5000 | D-5000 ③ | 0.6 | 30 | 0.2 | 10 | 5 | 7047 82 |
| ● 1000 - 10000 | D-10000 ④ | 0.6 | 60 | 0.2 | 20 | 10 | 7047 84 |



Transferpette® S, fixed volume

| Capacity, µl (color-coded) | Description | A* ≤ ± | | CV* ≤ | | Cat. No. |
|-------------------------------|-------------|--------|------|-------|------|----------------|
| | | % | µl | % | µl | |
| ● 10 | F-10 | 1 | 0.1 | 0.5 | 0.05 | 7047 08 |
| ● 20 | F-20** | 0.8 | 0.16 | 0.4 | 0.08 | 7047 16 |
| ● 25 | F-25 | 0.8 | 0.2 | 0.4 | 0.1 | 7047 20 |
| ● 50 | F-50 | 0.8 | 0.4 | 0.4 | 0.2 | 7047 28 |
| ● 100 | F-100 | 0.6 | 0.6 | 0.2 | 0.2 | 7047 38 |
| ● 200 | F-200** | 0.6 | 1.2 | 0.2 | 0.4 | 7047 44 |
| ● 500 | F-500 | 0.6 | 3 | 0.2 | 1 | 7047 54 |
| ● 1000 | F-1000 | 0.6 | 6 | 0.2 | 2 | 7047 62 |
| ● 2000 | F-2000 | 0.8 | 16 | 0.3 | 6 | 7047 64 |

* Calibrated to deliver (TD, Ex). Error limits according to the nominal capacity (= maximum volume) indicated on the instrument, obtained with instrument and distilled water at equilibrium with ambient temperature at 20 °C, and with smooth operation.
A = Accuracy, CV = Coefficient of variation

** For use only with 2-200 µl pipette tips

Note:

① D-1 for use only with nano-cap™ tip from BRAND.

② D-2.5: when using BRAND nano-cap™ tips or tips from other manufacturers, calibration and possibly adjustment is necessary.

③ D-5000 for use only with 5 ml tip from BRAND and 5 ml tip from Thermo Fisher Scientific Inc. (Finnpipette®).

④ D-10000 for use with 10 ml tip from BRAND and appropriate tips from Gilson® and Eppendorf®.



Pipette-Sets, Transferpette® S, adjustable volume

| Type | Description | Cat. No. |
|-------------------|---|----------------|
| MICRO | 3 Pipettes (D-2,5, D-10, D-100), 3 TipBox, 3 shelf/rack mounts | 7047 90 |
| MIDI | 3 Pipettes (D-20, D-200, D-1000), 3 TipBox, 3 shelf/rack mounts | 7047 91 |
| MACRO | 3 Pipettes (D-1000, D-5000, D-10000), 3 TipBox, 3 shelf/rack mounts | 7047 92 |
| STANDARD | 3 Pipettes (D-10, D-100, D-1000), 3 TipBox, 3 shelf/rack mounts | 7047 93 |
| Pipette package 1 | 5 Pipettes (D-2,5, D-10, D-100, D-200, D-1000), 3 TipBox, bench-top rack | 7048 50 |
| Pipette package 2 | 5 Pipettes (D-10, D-100, D-200, D-1000, D-5000), 4 TipBox, bench-top rack | 7048 51 |

Technical Data

Items supplied:

Transferpette® S-8/-12, DE-M marking, with performance certificate, 2 x TipBox, filled with pipette tips from BRAND, 1 shelf/rack mount, 1 reagent reservoir, 1 mounting tool, silicone grease and 1 set of sealing rings made of silicone/FKM.



Transferpette® S-8


| Capacity, µl (color-coded) | Description | A* ≤ ± % | CV* ≤ % | Subdivision µl | Cat. No. |
|-------------------------------|-------------|-------------|------------|-------------------|----------------|
| ● 0.5 - 10 | M8-10 | 1.6 | 1.0 | 0.01 | 7037 00 |
| ● 5 - 50 | M8-50 | 0.8 | 0.4 | 0.05 | 7037 06 |
| ● 10 - 100 | M8-100 | 0.8 | 0.3 | 0.1 | 7037 08 |
| ● 20 - 200 | M8-200 | 0.8 | 0.3 | 0.2 | 7037 10 |
| ● 30 - 300 | M8-300 | 0.6 | 0.3 | 0.5 | 7037 12 |



Transferpette® S-12

| Capacity, µl (color-coded) | Description | A* ≤ ± % | CV* ≤ % | Subdivision µl | Cat. No. |
|-------------------------------|-------------|-------------|------------|-------------------|----------------|
| ● 0.5 - 10 | M12-10 | 1.6 | 1.0 | 0.01 | 7037 20 |
| ● 5 - 50 | M12-50 | 0.8 | 0.4 | 0.05 | 7037 26 |
| ● 10 - 100 | M12-100 | 0.8 | 0.3 | 0.1 | 7037 28 |
| ● 20 - 200 | M12-200 | 0.8 | 0.3 | 0.2 | 7037 30 |
| ● 30 - 300 | M12-300 | 0.6 | 0.3 | 0.5 | 7037 32 |

* Calibrated to deliver (TD, Ex). Error limits according to the nominal capacity (= maximum volume) indicated on the instrument, obtained with instrument and distilled water at equilibrium with ambient temperature at 20 °C, and with smooth operation.
A = Accuracy, CV = Coefficient of variation

Note!  When ordering instruments with DAkKS calibration certificates, the prefix 'DAkKS' must be added to the order number, e.g., DAkKS 7047 28.

Accessories



Bench-top rack

For 6 Transferpette® S or Transferpette® S-8/-12 pipettes. Pack of 1.

Cat. No. **7048 05**



Shelf/rack mount

For 1 Transferpette® S single or multichannel pipette. Pack of 1.

Cat. No. **7048 10**



Individual stand

For 1 Transferpette® S pipette. Pack of 1.

Bis 1000 µl

Cat. No. **7053 85**

500-5000 µl

Cat. No. **7053 86**



Filter

Pack of 25.

| Description | Cat. No. |
|-------------|----------------|
| D-5000 | 7046 52 |
| D-10000 | 7046 53 |

For 1 Transferpette® S-8/-12 pipette. Pack of 1.

Cat. No. **7034 40**

The right pipette tip for your Transferpette *S*

Optimal performance will be achieved with the use of genuine premium tips from BRAND. BRAND quality tips seat perfectly for precise analyses.

DE-M

IVD

CE



| Pipette tips Filter tips Volume range | Transferpette® <i>S</i> single-channel Nominal volume | | | | | | | | | | | Transferpette® <i>S</i> multichannel Nominal volume | | | | | | |
|---|---|-----------------|-----------------|-------|-------|-------|--------|--------|--------|---------|------|---|-------|-------|-------|--------|--------|--------|
| | 1 µl | 2.5 µl | 10 µl | 20 µl | 25 µl | 50 µl | 100 µl | 200 µl | 500 µl | 1000 µl | 2 ml | 5 ml | 10 ml | 10 µl | 50 µl | 100 µl | 200 µl | 300 µl |
| | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● | ● |
| 0.1 - 20 µl | ✓ | ✓ ₂₎ | ✓ ₂₎ | | | | | | | | | | | ✓ | | | | |
| 0.5 - 20 µl | | ✓ | ✓ | | | | | | | | | | | ✓ | | | | |
| 1 - 50 µl | | ✓ ₂₎ | ✓ | | | | | | | | | | | ✓ | | | | |
| 2 - 200 µl ¹⁾ | | | | ✓ | ✓ | ✓ | ✓ | ✓ | | | | | | | ✓ | ✓ | ✓ | ✓ |
| 5 - 300 µl | | | | ✓ | ✓ | ✓ | ✓ | ✓ | | | | | | | ✓ | ✓ | ✓ | ✓ |
| 50 - 1000 µl ¹⁾ | | | | | | | | | ✓ | ✓ | | | | | | | | |
| 0.5 - 5 ml | | | | | | | | | | | ✓ | ✓ | | | | | | |
| 1 - 10 ml | | | | | | | | | | | | | ✓ | | | | | |
| 0,1 - 1 µl | ✓ | ✓ ₂₎ | ✓ ₂₎ | | | | | | | | | | | ✓ | | | | |
| 0.5 - 10 µl | | ✓ | ✓ | | | | | | | | | | | ✓ | | | | |
| 1 - 20 µl | | | ✓ | | | | | | | | | | | ✓ | | | | |
| 2 - 20 µl | | | | ✓ | ✓ | ✓ | ✓ | ✓ | | | | | | | ✓ | ✓ | ✓ | ✓ |
| 5 - 100 µl | | | | ✓ | ✓ | ✓ | ✓ | ✓ | | | | | | | ✓ | ✓ | ✓ | ✓ |
| 5 - 200 µl | | | | | | | | ✓ | | | | | | | ✓ | ✓ | ✓ | ✓ |
| 50 - 1000 µl | | | | | | | | | ✓ | ✓ | | | | | | | | |

✓ = Tip volume less than pipette's nominal volume

¹⁾ Bulk tips are yellow or blue colored, racked tips are colorless in a yellow-colored or blue-colored tip tray

²⁾ Calibration and possibly adjustment is necessary.

For more information about the Transferpette® *S* go to shop.brand.de!

BRAND®, Transferpette® and the BRAND word and figurative mark are registered trademarks of BRAND GMBH + CO KG, Germany. Other reproduced brands are the property of the respective owner.

Our technical literature is intended to inform and advise our customers. However, the validity of general empirical values, and of results obtained under test conditions, for specific applications depends on many factors beyond our control. Please appreciate, therefore, that no claims can be derived from our advice. The user is responsible for checking the appropriateness of the product for any particular application.

Subject to changes in design and equipment; errors and omissions excepted.

BRAND GMBH + CO KG · P.O. Box 1155 · 97861 Wertheim · Germany
Tel.: +49 9342 808-0 · Fax: +49 9342 808-98000 · E-Mail: info@brand.de · Internet: www.brand.de

